



Wearable multichannel haptic device for encoding proprioception in the upper limb

Downloaded from: <https://research.chalmers.se>, 2026-04-03 23:25 UTC

Citation for the original published paper (version of record):

Sagastegui Alva, P., Muceli, S., Atashzar, S. et al (2020). Wearable multichannel haptic device for encoding proprioception in the upper limb. *Journal of Neural Engineering*, 17(5).
<http://dx.doi.org/10.1088/1741-2552/aba6da>

N.B. When citing this work, cite the original published paper.

PAPER • OPEN ACCESS

Wearable multichannel haptic device for encoding proprioception in the upper limb

To cite this article: Patrick G Sagastegui Alva *et al* 2020 *J. Neural Eng.* 17 056035

View the [article online](#) for updates and enhancements.



The Department of Bioengineering at the University of Pittsburgh Swanson School of Engineering invites applications from accomplished individuals with a PhD or equivalent degree in bioengineering, biomedical engineering, or closely related disciplines for an open-rank, tenured/tenure-stream faculty position. We wish to recruit an individual with strong research accomplishments in Translational Bioengineering (i.e., leveraging basic science and engineering knowledge to develop innovative, translatable solutions impacting clinical practice and healthcare), with preference given to research focus on neuro-technologies, imaging, cardiovascular devices, and biomimetic and biorobotic design. It is expected that this individual will complement our current strengths in biomechanics, bioimaging, molecular, cellular, and systems engineering, medical product engineering, neural engineering, and tissue engineering and regenerative medicine. In addition, candidates must be committed to contributing to high quality education of a diverse student body at both the undergraduate and graduate levels.

[CLICK HERE FOR FURTHER DETAILS](#)

To ensure full consideration, applications must be received by June 30, 2019. However, applications will be reviewed as they are received. Early submission is highly encouraged.



PAPER

OPEN ACCESS

RECEIVED
10 May 2020REVISED
24 June 2020ACCEPTED FOR PUBLICATION
16 July 2020PUBLISHED
13 October 2020

Original Content from
this work may be used
under the terms of the
[Creative Commons
Attribution 4.0 licence](#).

Any further distribution
of this work must
maintain attribution to
the author(s) and the title
of the work, journal
citation and DOI.



Wearable multichannel haptic device for encoding proprioception in the upper limb

Patrick G Sagastegui Alva^{1,7} , Silvia Muceli^{2,7} , S Farokh Atashzar^{3,4,5,7} , Lucie William^{6,7}
and Dario Farina¹

¹ Department of Bioengineering, Imperial College London, London, United Kingdom

² Division of Signal Processing and Biomedical Engineering, Department of Electrical Engineering, Chalmers University of Technology, Gothenburg, Sweden

³ Department of Engineering, New York University, New York City (Brooklyn), NY, United States of America

⁴ Department of Mechanical and Aerospace Engineering, New York University, New York City (Brooklyn), NY, United States of America

⁵ NYU Wireless Center, 370 Jay Street, 9th Fl, Brooklyn, NY 11201, United States of America

⁶ INRIA, University of Montpellier, CNRS, Montpellier, France

E-mail: d.farina@imperial.ac.uk

Keywords: sensory substitution, vibrotactile feedback, haptics, assistive devices, biofeedback, prosthetic devices, wearable technologies

Abstract

Objective. We present the design, implementation, and evaluation of a wearable multichannel haptic system. The device is a wireless closed-loop armband driven by surface electromyography (EMG) and provides sensory feedback encoding proprioception. The study is motivated by restoring proprioception information in upper limb prostheses. **Approach.** The armband comprises eight vibrotactile actuators that generate distributed patterns of mechanical waves around the limb to stimulate perception and to transfer proportional information on the arm motion. An experimental study was conducted to assess: the sensory threshold in eight locations around the forearm, the user adaptation to the sensation provided by the device, the user performance in discriminating multiple stimulation levels, and the device performance in coding proprioception using four spatial patterns of stimulation. Eight able-bodied individuals performed reaching tasks by controlling a cursor with an EMG interface in a virtual environment. Vibrotactile patterns were tested with and without visual information on the cursor position with the addition of a random rotation of the reference control system to disturb the natural control and proprioception. **Main results.** The sensation threshold depended on the actuator position and increased over time. The maximum resolution for stimuli discrimination was four. Using this resolution, four patterns of vibrotactile activation with different spatial and magnitude properties were generated to evaluate their performance in enhancing proprioception. The optimal vibration pattern varied among the participants. When the feedback was used in closed-loop control with the EMG interface, the task success rate, completion time, execution efficiency, and average target-cursor distance improved for the optimal stimulation pattern compared to the condition without visual or haptic information on the cursor position. **Significance.** The results indicate that the vibrotactile device enhanced the participants' perceptual ability, suggesting that the proposed closed-loop system has the potential to code proprioception and enhance user performance in the presence of perceptual perturbation.

1. Introduction

Proprioception is an imperative sensory modality that is required for the proper execution of a wide range of activities of daily living. In the absence of a

biological limb, corresponding sensory information is partially lost along with motor abilities.

Myoelectric prostheses, which are controlled by electromyography (EMG) signals from the residual muscles of the stump [1], can partially replace the lost motor functionality using pattern decoding and machine learning. Although the available prostheses yield complex movements [2], they do not provide explicit sensory feedback. This degrades

⁷ These authors contributed equally and shared the first authorship

the functionality of the system and results in an unintuitive control [3–5]. Lack of sensory feedback results in excessive visual and cognitive loads for the users [3] and has been indicated as one of the main reasons causing a high rejection rate of active prostheses [5, 6].

Proprioception and haptics (sense of touch) are likely the two sensory modalities that are needed for prosthetic users [7]. Proprioception is the most challenging perception to restore because of its multimodal nature. Here, we focus on the restoration of proprioception using a wearable multichannel vibrotactile feedback system that provides a wide variety of spatiotemporal vibrotactile patterns.

The concepts of sensory supplementation and proprioception restoration have been investigated in past studies that relied on invasive or non-invasive approaches. Invasive methods use electrodes to directly stimulate the nerves [8, 9] or the primary sensory cortex in the brain [10]. For the non-invasive methods, cutaneous electrotactile [11] and vibrotactile [12] feedback systems have been tested in the literature, eliciting different sensations by activating different receptors on the skin [13]. To augment the bandwidth and complexity of the feedback, a typical approach is to increase the number of feedback channels [10]. These methods provide a sensation different from natural proprioception (sensory substitution). This information is provided by the modulation of parameters on the stimulation point, such as amplitude or frequency.

In this study, we focus on vibrotactile stimulation. The feasibility of using vibrotactile devices as sources of sensory information has been previously studied [14, 15], mainly for haptic sensation [16, 17]. For example, adding vibrotactile feedback for the touch sensation to myoelectric prostheses has shown improvement for task completion time and the controllability over force generation during grasping [18, 19]. Most previous studies have focused on grasp parameters, such as force control [18, 19], variability and characteristics of contact point [20], and the combination of force and one degree of freedom of hand motion [21]. Conversely, in this study, we investigate the encoding of multiple degrees of freedom of proprioception through multichannel closed-loop stimulation. This is achieved with an armband comprising eight vibrators. With this system and a myocontrol interface, we evaluated the ability of users to utilize the provided sensory feedback to complete a reaching task without visual feedback and in the presence of unpredictable perturbations of the reference control system.

It should be noted that the proprioceptive feedback provided by the system to the user was calculated in a closed-loop fashion based on the predicted intention of the user for generating the needed motion. In the experiments, we have also disturbed (by unpredictable random rotation) the mapping between the

natural motion and how the cursor moves. This was done to disturb the natural proprioceptive awareness of the user and thus to investigate whether the provided multichannel positional sensory feedback could recover the distracted proprioceptive awareness. The result of this analysis will illustrate the potential capability of the proposed system in restoring the lost proprioception. Thus, the proposed system provides information about the spatial motion of the user's limb to give the user concurrent awareness of where the limb position is in space so that the user can steer the control consequently. The distributed pattern of vibration provides positional information in parallel with the user's natural proprioception.

2. Materials and methods

2.1. Design of multichannel vibrotactile interface


The proposed closed-loop vibrotactile interface consists of a wearable wireless device (figure 1) that allows full customization of closed-loop stimulation patterns. This wearable armband embeds eight vibration modules. Each module includes a motor (310-103 Vibration ERM, Precision Microdrive's, USA), and the driver (DR2605L Haptic Driver, Texas Instruments, USA). An embedded Arduino Micro was used to generate the control signals of the waveforms for the vibration patterns.

In order to control the motors and provide the needed voltage, each stimulating module includes embedded amplifiers. The motors' driver was a DRV2605L to play custom-made waveforms directly through the integrated circuit (I2C). This allowed us to control all motors by using only four wires, which would allow for the expansion of vibration units on the same I2C bus. In addition, an address translator (LTC4316, Analog Devices, USA) was utilized to modify the default address of the motor driver.

The frequency of the vibrotactile motors was controlled by tuning the voltage to each motor separately. The voltage was computed by the microprocessor and sent by I2C to the corresponding motor driver. An increase of the voltage determined a simultaneous increase of the vibration amplitude and frequency, due to the coupled characteristics of the vibrotactile motors. The communication between the PC and the armband occurred via USB and Bluetooth v2.0.

2.2. Demographic data

The study included two experimental sessions run in two days. Nine able-bodied participants were included in the experiments. The experimental procedure was approved by the Imperial College Research Ethics Committee (ref number 18IC4685). All participants signed an informed consent before the experiments. Experiment 1 (four males and four females,



age 27.0 ± 3.1 years) aimed at determining the sensitivity to vibration around the forearm delivered by the developed armband and the ability to discriminate different levels of vibrations. Experiment 2 (five males and three females, age 27.0 ± 3.1 years) aimed at investigating if proprioception information could be substituted by the vibration. Seven subjects were common to both experiments.

2.3. Experimental setup

The experimental setup comprised:

- A host PC (Intel® Core(TM) i7-8750 H CPU at 2.20 GHz, 16 GB RAM) running a closed-loop myocontrol framework and an external 15.6" screen to provide visual feedback to the participants
- Sixteen pre-gelled monopolar EMG electrodes (Ambu Neuroline 72 000-S/25, Denmark)
- A ground electrode

- An EMG amplifier (OT EMG USB2+, OT Bioelettronica, Italy)
- The developed multichannel vibrotactile interface (section 2.1).

The vibrotactile interface (armband) was placed on the proximal third of the forearm, according to figure 1(D). Using the specific mechanical design, motors around the forearm were positioned so that motor #1 was on top of the ulna, and motor #5 was at the opposite side. The other six motors were placed equidistant, with three motors (motor #2, #3, and #4) on the volar side and the others on the dorsal side (see figure 1(D)). Sixteen EMG electrodes were placed in two rows circumferentially and equidistant around the forearm, distally with respect to the armband to record muscle electrical activity. The center-to-center electrode distance was 2 cm in the forearm longitudinal direction. The center-to-center distance between the motors and the most proximal row of electrodes was 2.5 cm. The ground electrode was placed at the wrist. A table-top EMG amplifier, OT EMG USB2+, was used to retrieve high-quality signals for reliable control. The amplification gain was set to be 100 V/V, sample rate at 2048 Hz with a band-pass filter at 10–500 Hz. The control loop was implemented using a myoelectric interface developed in MATLAB that processed and decoded the EMG signals. The interface displayed a virtual environment that represented the position of the hand (prosthesis).

Two degrees of freedom of the wrist were used for online control (wrist flexion/extension and radial/ulnar deviation). A linear regressor was used as the machine learning core of the implemented software to map the EMG signals to cursor movements. The input for the linear regressor was the RMS of the EMG signals calculated over sliding time windows of 160 ms with 120 ms overlap. Three series of movements were included in the training phase: flexion/extension, radial/ulnar deviation, and a combination of the two along with the 45° , 135° , 225° , and 315° directions. Each movement was repeated three times. The trained regressor was used in the controller allowing for simultaneous and proportional steering of the two degrees of freedom. There were no visible artefacts due to the stimulation on the EMG; the cursor control was not affected when the vibration was activated. This can also be concluded from the high performance in the task when the visual feedback was provided (the simplest motor control task), in which the system was able to regress the intended motion of the user.

2.4. Experiment 1: Resolution

Experiment 1 was designed to determine the individual's sensitivity to the vibration system and the

ability to discriminate different stimulation parameters. Participants were comfortably seated in front of a table with the dominant (right) arm in a rest position (palm facing inwards and thumb pointing upwards) over an arm support. Experiment 1 included an initial assessment of the sensation threshold (Experiment 1.a), a vibration intensity discrimination test (Experiment 1.b), and a final re-assessment of the sensation threshold (Experiment 1.c).

In Experiment 1.a, the vibration frequency was incremented from 0 Hz (no stimulation) in steps of 3% (6.69 Hz) of the motor vibration ($V_{\max} = 232$ Hz) until the participant reported perceiving the stimulus (V_{12h_m} , where m is the motor index). Then, the vibration frequency was set to 40% of the maximal vibration of the motors (116.4 Hz) and was decremented until the participant reported feeling no sensation of the vibration (V_{h2l_m}). For each motor, we considered as sensation threshold ($V_{th_m,i}$, where i stands for initial) the maximum between V_{12h_m} and V_{h2l_m} . The lower limit (V_{L_m}) used during Experiment 1.b was $1.1 \cdot V_{th_m,i}$.

In Experiment 1.b, we determined the number of vibrotactile feedback levels that the user could effectively discriminate. We refer to this test as intensity discrimination rather than frequency discrimination because vibration frequency and amplitude are coupled in the motors we used (section 2.1). The provided levels of vibration ranged from V_{L_m} to V_{\max} . The vibration range was discretized in four and five equidistant levels. The different numbers of levels were provided in a randomized order across different participants. As in previous studies [22, 23], for these tests, we included learning, reinforced learning, and validation. The learning phase was meant to familiarize the participant with the vibration levels. For each location, a series of different vibration levels was presented five times in ascending order, and the level delivered was shown on a screen in front of the participant. For the reinforced learning phase, a random stimulus of the four or five levels was delivered to the participant until the perceived level was reported and selected on the interface. The correct answer was then shown on the screen. All levels were presented randomly five times each (20 trials for 4 levels and 25 trials for 5 levels). The validation step had a structure similar to the reinforced learning step, but in this case, the answer was not provided.

Experiment 1.c aimed at investigating the adaptation phenomena, an increase in the sensation threshold after a persisting stimulation by vibration. This phenomenon has been reported in the literature and has a fast recovery process for the vibrotactile sensation (2 or 3 min) [24]. Following Experiment 1.b, the sensation threshold for the position corresponding to each motor ($V_{th_m,f}$, where f stands for final) was determined again following the procedure described for Experiment 1.a.




Figure 2. Virtual reality environment for position control. Flexion/extension movements controlled the horizontal direction of the red cursor, ulnar/radial deviation controlled its vertical direction. The magnitude of the movement is the distance to the rest position (the center of the environment), M , and the angular position is given by α . The purple circle is the target. The cyan circle, with radius r , is the workspace where the target may appear.

2.5. Experiment 2: Evaluation of proprioception encoding

The reliability and performance to control the position of the hand in closed-loop were evaluated in Experiment 2 through the delivery of different feedback patterns around the forearm to map the direction of the wrist. The task consisted of reaching 20 random targets in a virtual environment (figure 2).

As in Experiment 1, participants were comfortably seated in front of a table with the dominant arm over a support, the hand palm facing inwards, and thumb pointing upwards. The position of the user's hand was mapped to the motion of a red cursor in the virtual representation. Wrist extension and flexion movements (when detected correctly by the machine learning algorithm) produced displacements of the cursor to the right and left, respectively, whereas radial and ulnar deviation moved the cursor respectively up and down. During the test phase, the mapping was perturbed by applying a random rotation of $+90$ or -90 degrees. This rotation was introduced to investigate whether participants could rely more on the added sensory channels (in comparison to their natural proprioception) to reconstruct the perturbed sensory navigation task.

Three feedback modes were evaluated to perform the task: *Visual*, *NVNV* (no visual no vibration), and *Vibrotactile* feedback. During the *Visual* feedback modality, the participant could see the cursor on the screen. For the other feedback modalities (*Vibrotactile* and *NVNV* mode), the cursor was hidden so that the participants could only see when the cursor hit the target at the end of the task by changing the color to green if it is hit correctly. In the case of the *Vibrotactile* feedback, four patterns of vibration were tested, as

















Feedback Modality	1st Level	2nd Level	3rd Level	4th Level
<i>Visual</i>	Continuous mapping			
<i>NVNV</i>	No Mapping			
<i>VibSpMod</i>				
<i>VibFrMod3-3</i>				
<i>VibFrMod1-3</i>				
<i>VibFrMod1-1</i>				

Figure 3. Different feedback modalities. In the code for the *Vibrotactile* modality, *SpMod* stands for spatial modulation, *FrMod* for frequency modulation, the first number indicates the *number of modulated motors*, and the second the *number of active motors*. *NVNV* stands for no vision, no vibration. The intensity of the colors (light vs dark) corresponds to the intensity of the vibration at the corresponding point (low vs high intensity).

described below. No vibration and no visual feedback were applied in the *NVNV* mode. The task was considered successful if the participant was able to position the red cursor within the purple circle for at least 300 ms (dwelling time) consecutively, within 20 s (task failure time) [25]. In this case, the color of the purple circle changed to green to notify the user that the task was successful.

The regressor processed the 16 surface EMG input signals and estimated the intended intensity and orientation of the motion. The estimated intended motion was used to control the two characteristics of the patterns of vibrotactile feedback (i.e. the intensity and direction of vibration) to close the loop. The sensory feedback was provided as follows: the amplitude of the movement (M) was mapped to the number of active motors and to the frequency of vibration of the active motors while the angle (α) was associated to the location of the group of active motors. The whole range of amplitude of the movements was divided into four levels of vibration frequency (see section 3.1).

The patterns used to map the proprioception are shown in figure 3. The green color represents the active motors, and the frequency of vibration is represented by the color intensity with darker colors corresponding to a higher frequency. The control performance obtained with these patterns were compared to the *Visual* and *NVNV* modes that acted as a benchmark.

In figure 3, *VibSpMod* refers to one possible activation pattern that started with one motor aligned with the direction of the movement and then it activated the surrounding motors at the same frequency while the number of motors activated increased with the

amplitude of the movement (level). *VibAmpMod3-3* activated three motors, while their frequency increased with the level. For *VibAmpMod1-1* only one motor was used to map the cursor direction and the frequency increased with the level. *VibAmpMod1-3* activated always three motors, but only the frequency of the central one increased with the level.

For each feedback pattern, the user had 2 min to free play with both the *Visual* and *Vibrotactile* feedback modalities in order to explore the different stimulation levels around the workspace (for the *NVNV* feedback, as there was no feedback at all, participants could not explore the environment). The mapping rotation was not applied during the familiarization phase. Then the user had the task to reach the 20 targets based only on the vibrotactile feedback, the visual feedback or no feedback (*NVNV* mode). The red cursor that indicated the visual online feedback was disabled for all patterns except for the *Visual* feedback.

2.6. Data analysis

For Experiment 1.b, the outcome measure was the success rate, expressed as the percentage of the correctly identified vibration levels. Confusion matrices were also generated to evaluate and visualize how much the participants confused different levels and how the confusion varied among different levels. For Experiments 1.a and 1.c, one-way ANOVA and Turkey's honestly significant criterion were used to compare the sensation threshold in different physiological positions (i.e. corresponding to different motors) and the initial ($V_{th_m_i}$) and final ($V_{th_m_f}$) thresholds. For Experiment 1.b, one-way ANOVA was used to compare the ability of participants to distinguish four or five vibration levels. For Experiment 2, the following performance metrics were evaluated: (a) the percentage of reached targets out of the 20 attempted (success rate), (b) the average distance between the cursor's position and the target along the traveled path, (c) the median pathway efficiency (PathE), (d) the median duration of the trial (trial time), (e) the median speed, (f) the median completion time of the successful trials (successful trial time).

The PathE was defined as the ratio between the shortest pathway to reach the target and the path traveled by the cursor, expressed as a percentage. The metrics used for quantifying the trial duration (trial time and successful trial time) were complementary. For example, trial time discriminated between a participant who quickly reached a target once and failed all the other trials and a participant who reached all targets in a similar time. Successful trial time discriminated a participant who could accomplish a reach in >20 s from another one who could not finish irrespective of the time given.

As data did not pass the normality Shapiro test, the non-parametric Wilcoxon signed-rank test was used. The threshold for the statistical significance

was set at $p < 0.05$. The results were expressed as *mean \pm standard deviation* or as a box plot.

3. Results

3.1. Experiment 1: resolution

The statistical analysis demonstrated a significant difference of the sensation threshold for all participants and all the motors around the forearm between Experiment 1.a (initial, 87.2 ± 12.6 Hz) and Experiment 1.c (final, 100.1 ± 15.3 Hz, $p < 0.001$, see figure 4(A)).

In addition, a significant difference was found between the initial (Experiment 1.a) and final (Experiment 1.c) sensation threshold for all locations (motors), except for the ulnar and radial positions (see black starts in figure 4(B), physiological positions 1 and 5).

Statistical differences were observed between the sensation thresholds of the different physiological positions of stimulation (both Experiment 1.a (green stars) and 1.c (blue stars), see figure 4(B)). In all cases, the threshold of the forearm region situated on top of the ulna (physiological position 1) was significantly lower than the threshold of the dorsal side of the forearm (physiological positions 6, 7 and 8). The statistical analysis of the results for Experiment 1.b demonstrated a significant difference in the success rate for all participants and all motors around the forearm for the discrimination of 4 and 5 levels of intensity ($86.6 \pm 11.4\%$ and $75.7 \pm 11.5\%$, respectively $p < 0.001$, see figure 5(A)). Additionally, on average around the forearm, there were significant statistical differences between the success rate for 4 and 5 different levels of intensity for the positions 2, 5, 6 and 7 (corresponding to the ulnar/volar, the radial, the radial/dorsal and the dorsal parts of the forearm, see 5, (B) and (C)).

3.2. Experiment 2: Evaluation of proprioception encoding

Experiment 2 tested the ability to perceive the cursor movement with different feedback strategies (*Visual* mode, *NVNV*, *vibrotactile* mode).

Figure 6 shows the performance of each participant (column) for each performance metrics (rows A–F) and feedback modalities. In each panel, the green area corresponds to the region where the results are better than *NVNV* feedback, and the grey area is the one where the results are worse than *NVNV* feedback.

Figure 7 shows a representative example of differences between the path traveled in case of the *NVNV* feedback and the best feedback (in that case, *VibFrMod3-3* feedback) for one participant.

From the results of figure 6, it is possible to find for each subject the best vibrotactile pattern which had the most frequent statistical significant difference or higher success rate. For each individual, such

a pattern is the one that has the maximum benefit for the user in terms of substituting proprioception when compared with no feedback experiment (*NVNV* mode).

The feedback with the best performance was *VibFrMod1-1* for participants 5 and 8, *VibFrMod1-3* for participants 3 and 6, *VibFrMod3-3* for participants 2 and 7 and *VibSpMod* for participants 1 and 4.

As expected, *Visual* feedback resulted in better performance compared to *NVNV* (figure 8). In fact, success rate and path efficiency were higher, the average distance from the target and the time to accomplish the trials and successful trials were lower. However, speed was lower in case of *Visual* feedback.

The comparison of the *Visual* and *NVNV* feedback with the *Best* vibrotactile feedback demonstrated an improvement of the performance with the vibrotactile feedback compared to the *NVNV* mode for all metrics but speed, and no statistical differences with respect to the *Visual* feedback for 3 (success rate, path efficiency, successful trial time) out of the 6 considered metrics (figure 8).


4. Discussion

In this study, a multichannel wearable vibrotactile device was proposed to encode proprioceptive information. We showed that this could be achieved using eight channels of vibrotactile feedback spatially distributed around the arm and four vibration levels per channel controlled in a closed-loop manner based on measurements of surface EMG. The custom-made system features a modular design that supports straightforward modification of the number of vibration units and complete control of the vibration characteristics.

4.1. Experiment 1: resolution

Experiment 1 provided a psychometric characterization of the proposed vibration armband. Experiment 1.a and 1.c showed differences between the threshold perceived in different positions around the forearm. The sensation thresholds of the ulnar and volar/ulnar portions of the forearm were lower than those of the dorsal region. Therefore, the observed difference in the vibration sensation threshold in different forearm positions may be due to anatomy, e.g. differences in the density of the Pacinian corpuscles (mechanoreceptors responsible for vibration) in the dermis, or the dermis thickness where Pacinian corpuscles are located [25].

The relatively high threshold frequency is determined by the simultaneous modulation of amplitude and frequency with the voltage of the motor. The amplitude of the motor has a steeper slope with the voltage than the frequency. Consequently, at low voltage and frequency, the amplitude is considerably low. An increase of the sensation threshold was observed between Experiment 1.a and Experiment




1.c, which could be associated with sensory adaptation. For a prolonged use of the device, the observed adaptation could be taken into account by gradually increasing the stimulation intensity. In Experiment 1.b, participants were able to discriminate 4 levels of vibration intensity for each motor better than 5 (figure 5(A)). Classification accuracy of 86.6% was obtained in discriminating four vibration levels compared to 75.7% in the case of five levels (figures 5(D) and (E)). Likely, accuracy would have further improved in case of three vibration levels only. However, in Experiment 2 we decided to maintain four levels for the various vibration patterns to convey high-resolution information about proprioception.

4.2. Experiment 2: evaluation of proprioception encoding

In Experiment 2, four patterns of vibration (shown in figure 3) with variable location and distribution of stimulation intensity were tested in comparison with two modes where the participants could rely on visual feedback (*Visual* mode) or were deprived of both visual and vibrotactile feedback (*NVNV* mode). The *Visual* mode was provided with the most comprehensive source of feedback. Therefore,

all metrics were better compared to the *NVNV* mode where feedback was absent, with the only exception of speed (figure 8). Participants were faster in the *NVNV* mode. This was likely due to the fact they started moving very fast around the workspace until they accidentally hit the target. This is also reflected in the result that the path taken was quite long and random in the *NVNV* mode (figure 7(A)).

When provided with vibrotactile feedback (*Best* pattern), performance improved with respect to the *NVNV* mode. For each subject, we selected the pattern that provided the greatest improvement compared to the absence of feedback. Each vibrotactile pattern resulted to be the best for 25% of the participants (figure 6). We believe that the underlying mechanism which causes one pattern to be the best for one particular subject depends on the neurophysiological differences between the users. For amputees, the effectiveness of a specific stimulation pattern with respect to others will be affected by the type of surgery and potential nerve and muscle damages during the surgery. Therefore, the optimal pattern will need to be determined on a patient-specific basis. Although the test of multiple patterns may be cumbersome




in patients, the initial evaluation of the optimal stimulation pattern will need to be done only once at the time of prosthetic fitting and revised occasionally.

The *Best* pattern among the four patterns of vibration tested in Experiment 2 determined an improvement of five out of the six performance metrics compared to the *NVNV* mode (figure 8). Only speed was worst, likely for the same reason mentioned in *Visual Mode*. It is relevant to note that speed is a measure of performance which should be interpreted together with other measures, such as accuracy or success rate. In our experiments, when users did not have any feedback, they started random fast motions exploring the whole workspace with the hope of hitting the target in the given time. This random motion resulted in a higher speed which was unrelated to accuracy in control. Conversely, when the user tracked the target by using sensory feedback, the increase in speed corresponded to better control, as revealed by the other performance metrics. Thus, high speed can represent opposing scenarios in terms of quality of the control. As a result, speed should not be studied as a stand-alone measure. However, alongside other measures which can decode the ability of the user in interpreting the sensory feedback resulting in high motor performance, higher speed

can be considered as an indicator for improved motor ability. In three out of six metrics, the vibrotactile feedback reached a similar performance as the *Visual* mode (figure 8).



Since able-bodied subjects, differently from amputees, have intact proprioception, a sensory perturbation (in the form of random rotation of the control) was applied to ensure that participants actually relied on the sensory feedback provided by the vibration rather than on their own proprioception. This was evident in the path taken, as shown in the representative example in figure 7(B). When the target appeared, the device wearer started to contract the proper muscles in order to reach the target. However, this resulted in an diverging undesired movement due to the random rotation applied by the system. Therefore, he/she had to change strategy and rely on the cue provided by the vibration to accomplish the reaching task and control successfully. In fact, figure 7(B) shows the trajectory was directed toward the target once the user started relying on the vibrotactile feedback after an initial divergence due to the rotation in control. This initial divergence resulted in a decrease of path efficiency compared to previous studies where a similar myoelectric control paradigm was applied but the control was unperturbed [25]. The performance metrics found when the vibrotactile stimulation



was applied have to be considered in relation to the NVNV and *Visual* modes rather than in absolute terms due to the paradigm employed (proprioception perturbation) in order to evaluate the performance of the proposed closed-loop system when sensation is impaired. Taken together, the results showed that the participants were able to exploit the proprioceptive information conveyed by the proposed vibrotactile device to overcome the perturbation imposed by the control system. We can conclude that the proposed

device showed significant flexibility in controlling the stimulation parameters and reconstructing the perturbed proprioception.

With the vibrational feedback, the user was able to recover the performance which was comparable with that of visual feedback for some of the outcome measures. One of the major problems with the use of current prosthetic devices is that the user needs to visually track the motions of the limb to be able to successfully implement a meaningful motion. This results



in high visual processing load during task execution, which negatively affects the usability of prostheses. This is one of the reasons for rejection of active prosthetic devices by many users. An ideal substitutional proprioceptive feedback could reduce the need for visual inspection of motor tasks. This study provides a step forward in this direction. We have shown that by the particular spatial distribution of vibrotactile feedback and through modulation of the intensity of the feedback, we were able to recover several aspects of motor control. This shows the importance and significance of vibrotactile feedback and the corresponding potential for recovering the lost proprioception. With the use of the proposed armband, the user will also have the option of combining visual and vibrotactile feedback, which may result in a multimodal approach to compensate for the lost proprioception.

In this study, the proposed device was systematically tested in able-bodied users to provide a

benchmark for future implementation in amputees. However, the paradigm was motivated and designed in order to simulate the proprioception impairment occurring in individuals with limb deficiency. Our future step will be to test the device in amputees to investigate whether axonal regrowth or phantom limb can affect the performance of the users in utilizing substitutional sensory channels for recovering proprioception.

5. Conclusion

In this paper, we have systematically evaluated the performance of a surface EMG-driven closed-loop sensory feedback system based on a wearable vibrotactile haptic device that provides eight channels of vibrotactile feedback for the encoding of proprioception. The results strongly support the capability of the proposed device of substituting the perturbed proprioception, suggesting a customizable pattern of


stimulation with user-specific level of vibration for each participant. The next step will be to investigate the use of the device and test patterns with amputees.

6. Funding

This project was funded by the European Research Council (ERC) under the European Union's Horizon 2020 research and innovation programme (project NaturalBionicS; Grant Agreement No. 810 346). SM was financed by the Life Science Engineering Area of Advance at Chalmers University of Technology.


ORCID iDs

Patrick G Sagastegui Alva  <https://orcid.org/0000-0001-7087-7789>

Silvia Muceli  <https://orcid.org/0000-0002-0310-1021>

S Farokh Atashzar  <https://orcid.org/0000-0001-8495-8440>

Lucie William  <https://orcid.org/0000-0002-1590-5211>

Dario Farina  <https://orcid.org/0000-0002-7883-2697>

References

- [1] Farina D, Jiang N, Rehbaum H, Holobar A, Graimann B, Dietl H and Aszmann O C 2014 The extraction of neural information from the surface EMG for the control of upper-limb prostheses: emerging avenues and challenges *IEEE Trans. Neural Syst. Rehabil. Eng.* **22** 797–809
- [2] Belter J T, Segil J L, Dollar A M and Weir R F 2013 Mechanical design and performance specifications of anthropomorphic prosthetic hands: A review *J. Rehabil. Res. Dev.* **50** 599–618
- [3] Pylatiuk C and Döderlein L 2007 Results of an internet survey of myoelectric prosthetic hand users *Prosthet. Orthot. Int.* **31** 362–70
- [4] Jang C H, Yang H S, Yang H E, Lee S Y, Kwon J W, Yun B D, Choi J Y, Kim S N and Jeong H W 2011 A survey on activities of daily living and occupations of upper extremity amputees *Annals Rehabil. Med.* **35** 907–21
- [5] Cordella F, Ciancio A L, Sacchetti R, Davalli A, Cutti A G, Guglielmelli E and Zollo L 2016 Literature review on needs of upper limb prosthesis users *Front. Neurosci.* **10** 209
- [6] Lewis S, Russold M, Dietl H and Kaniusas E 2012 User demands for sensory feedback in upper extremity prostheses 2012 *Int. Symp. on Medical Measurements and Appl. Proc.* pp 1–4
- [7] Antfolk C, D'alozzo M, Rosén B, Lundborg G, Sebelius F and Cipriani C 2013 Sensory feedback in upper limb prosthetics *Expert Rev. Med. Dev.* **10** 45–54
- [8] D'Anna E et al 2019 A closed-loop hand prosthesis with simultaneous intraneural tactile and position feedback *Sci. Robotics* **4** 88–92
- [9] Dhillon G S and Horch K W 2005 Direct neural sensory feedback and control of a prosthetic arm *IEEE Trans. Neural Systems Rehabil. Eng.* **13** 468–72
- [10] Hsiao S S, Fettiplace M and Darbandi B 2011 Sensory feedback for upper limb prostheses *Progress in Brain Research* (Amsterdam: Elsevier) vol 192 pp 69–81
- [11] Dosen S, Markovic M, Štrbac M, Belić M, Kojić V, Bijelić G, Keller T and Farina D 2017 Multichannel electro-tactile feedback with spatial and mixed coding for closed-loop control of grasping force in hand prostheses *IEEE Trans. Neural Syst. Rehabil. Eng.* **25** 183–95
- [12] Ko C Y, Chang Y, Kim S B, Kim S, Kim G, Ryu J and Mun M 2015 Evaluation of physical and emotional responses to vibrotactile stimulation of the forearm in young adults, the elderly and transradial amputees *Physiol. Behavior* **138** 87–93
- [13] Kaczmarek K A, Webster J G, Bach-y-Rita P and Tompkins W J 1991 Electrotactile and vibrotactile displays for sensory substitution systems *IEEE Trans. Biomed. Eng.* **38** 1–16
- [14] Oakley I, Kim Y, Lee J and Ryu J 2006 Determining the feasibility of forearm mounted vibrotactile displays 2006 *14th Symp. on Haptic Interfaces for Virtual Environment and Teleoperator Systems* (IEEE) pp 27–34
- [15] Jones L A and Sarter N B 2008 Tactile displays: Guidance for their design and application *Hum. Factors* **50** 90–111
- [16] Witteveen H J, Droog E A, Rietman J S and Veltink P H 2012 Vibro-and electro-tactile user feedback on hand opening for myoelectric forearm prostheses *IEEE Trans. Biomed. Eng.* **59** 2219–26
- [17] Christiansen R, Contreras-Vidal J L, Gillespie R B, Shewokis P A and O'Malley M K 2013 Vibrotactile feedback of pose error enhances myoelectric control of a prosthetic hand *World Haptics Conf.* pp 531–6
- [18] De Nunzio A, Dosen S, Lemling S, Marković M, Schweisfurth M, Ge N, Graimann B, Falla D and Farina D 2017 Tactile feedback is an effective instrument for the training of grasping with a prosthesis at low- and medium-force levels *Exp. Brain Res.* **235** 2547–59
- [19] Clemente F, D'Alonzo M, Controzzi M, Edin B B and Cipriani C 2016 Non-invasive, temporally discrete feedback of object contact and release improves grasp control of closed-loop myoelectric transradial prostheses *IEEE Trans. Neural Syst. Rehabil. Eng.* **24** 1314–22
- [20] Ninu A, Dosen S, Muceli S, Rattay F, Dietl H and Farina D 2014 Closed-loop control of grasping with a myoelectric hand prosthesis: Which are the relevant feedback variables for force control? *IEEE Trans. Neural Syst. Rehabil. Eng.* **22** 1041–52
- [21] Witteveen H J B, Luft F, Rietman J S and Veltink P H 2014 Stiffness feedback for myoelectric forearm prostheses using vibrotactile stimulation *IEEE Trans. Neural Syst. Rehabil. Eng.* **22** 53–61
- [22] Štrbac M et al 2016 Integrated and flexible multichannel interface for electro-tactile stimulation *J. Neural Eng.* **13** 046014
- [23] Muceli S, Poppendieck W, Hoffmann K P, Dosen S, Benito-León J, Barroso F O, Pons J L and Farina D 2019 A thin-film multichannel electrode for muscle recording and stimulation in neuroprosthetics applications *J. Neural Eng.* **16** 026035
- [24] Barker D E 1951 Skin thickness in the human *Plastic Reconstructive Surgery* **7** 115–16
- [25] Muceli S, Jiang N and Farina D 2013 Extracting signals robust to electrode number and shift for online simultaneous and proportional myoelectric control by factorization algorithms *IEEE Trans. Rehabil. Eng.* **22** 623–33